



Arthrex, Inc.
 1370 Creekside Blvd.
 Naples, Florida 34108

Test Information

Specimen properties: Specimen label	PT Magz EZ Speed Loader
Specimen text inputs: Part	
Specimen text inputs: Lot	
Specimen text inputs: Operator	
Specimen text inputs: Test	Straight Pull
Specimen text inputs: Date	
Data: Interval 1	500.00000 Hz
Test: Rate 1	1.00000 in/min
Specimen text inputs: Instron Model	E10000, S/N: E10BMTU2901
Specimen text inputs: Load Cell Model	10 kN / 100 Nm, S/N: 124258 (Force) / 120687 (Torque)

Results Table

	Yield Load [lbf]	Maximum Load [lbf]	Mode of Failure
1	135.15	168.32	Telescoping tube broke
Mean	135.15	168.32	
Standard Deviation	-----	-----	

Graph

PT Magz EZ Speed Loader

